INSTRUCTIONS FOR USE

RYLE'S TUBE RYLES-CARE

Material: PVC

Instructions for use:

- To perform the procedure, instruct the patient to hold his head straight up with his neck slightly hyper extended and facing forward. Grasp the end of the tube above the lubricant. With the curve pointing downward, carefully insert the tube along the floor of the nostril, on the lateral side.
- Once the tube reaches the nasopharynx, you may meet some resistance. This is the point at which the patient may start to gag or cough. Have him tilt his head forward-unless medically contraindicated. This positioning will close the trachea and make it easier for the tube to pass through the esophagus.
- Offering your patient sips of water helps move the tube past the oropharynx. Slowly advance the tube each time he swallows. If he continues to cough or choke and is unable to talk, withdrawn the tube slightly since it may have accidently entered the trachea.
- If that doesn't help, look inside his mouth to see if he tube has become coiled at the back of the throat. If so, slowly withdraw the tube until it is straight and allow the patient to rest briefly before you try again.
- If several of your attempts are unsuccessful, stop and notify your supervisor or the attending physician. Chances are that will not be necessary, though and you will manage to pass the tube to the desired length. When the taped mark reaches the nostril, stop and confirm proper placement in the stomach.
- Re-use of this device may change its mechanical or biological features and may cause device failure, allergic reactions of infections.

Warning:

- To be use this product by a qualified doctor or a paramedic.
- Read instructions before use.

MANUFACTURER DISCLAIMS ANY RESPONSIBILITY FOR POSSIBLE CONSEQUENCES RESULTING FROM IMPROPER USE.

· The product should not be reprocessed.
· Visually inspect and carefully check the product and packaging before use.
· Improper transport and handling may cause structural and /or functional damage to device or packaging.
· The product is guaranteed non-toxic, sterile and Pyrogen free if the package is not opened or damaged.
· Do not clean or re-sterilize.
· The product should be used immediately after opening the packaging.
- The device must not be re-used after its usage, it con be a potential biohazard.

Storage Condition :

· Do not expose to heat or direct sunlight

CE 0123

MD Medical Device
REF Reference No.
LOT Lot No./Batch No.
Manufacturer
Date Of Manufacturing
Use By / Expiry Date
STERILE EO Sterile Barrier System Sterilised By Ethylene Oxide
Do Not Reuse
Do Not Resterilize
Non-Pyrogenic
Do Not Use If Package Is Damage
Caution
PHT Contains phthalate
Consult Instructions For Use
10°C 45°C Temperature Limit
EC REP Authorised Representative In The European Community

Disposafe Health and Life Care Limited
Plot No. 1 & 2, Phase-II, Sector-59, Ballabhgarh,
Faridabad-121004, Haryana, (INDIA)
E-mail : info@disposafe.com,
Web : www.disposafe.com
Customer Care No. 91-129-4170201

EC REP CEpartner4U BV, ESDOORNLAAN 13, 3951 DB MAARN, THE NETHERLANDS.
E-mail : office@cepartner4u.com

Mfg. Lic. No.: MFG/MD/2019/000106
AW/IFU_790, REV 02, Dt. 03-03-2023